



Data treatment

Pål Rustad
NORIP workshop
Reykjavik, 10/8-2002



Controls

- 5 control samples:
 - CAL: Serum pool, target values for most components established by reference methods
 - X: Serum pool from men
 - P: Serum pool from women using contraceptive pills
 - HIGH: Serum pool concentrated by freezing
 - LOW: HIGH diluted 1:2 with calcium/sodium-solution





Analysis

Person samples measured together with control samples is defined as one series.

- At least one series should consist of:
 - 10 CAL
 - 3 each of X, P, HIGH, LOW
- other series by:
 - 10 X

Selection/calculation of reference values





Calculation

except enzymes

- For all components (except enzymes) CAL is the reference standard (calibrator), i.e. all results (with CAL in series) are multiplied by the factor:

$$\text{CAL}_{\text{target}} / \text{CAL}_{\text{mean of this series}}$$


- Factor for series with only X:

$$X_{\text{target}} / X_{\text{mean of this series}}$$

where

$$X_{\text{target}} = \text{CAL}_{\text{target}} * \text{median}(X_{\text{mean}} / \text{CAL}_{\text{mean}})$$

- Results: Creat, etc



Calculation enzymes

- R - result from lab
- r - result from instrument
- S, I - slope, intercept used by lab as correction parameters according to formula:

$$R = S \cdot r + I \text{ or } r = (R-I)/S$$

r is converted to common unit (U/L)



Error handling

- Laboratory asked to check registered data by printing report from registration program before submitting data!
- Report
- Person analytical data
 - comparison of different materials for same person and component
- Control data



Exclusion of data

- According table with person analytical data, these are the concepts of exclusion:
 - Person (Peter Feldings)
 - Enzymes (Heidi Steenslands)
 - Controls
 - Material (comparison of results from fresh and thawed serum and plasma for same person)
 - Gross errors (missing decimal point, etc)
 - Duplicate



Exclusion of data

Controls

- Quality goals for precision NOT evaluated
- Quality goals for systematic error (bias):
 - optimum: $B < 1/8 \cdot s_B + k \cdot s_A$
 - desirable: $B < 1/4 \cdot s_B + k \cdot s_A$
 - minimum: $B < 3/8 \cdot s_B + k \cdot s_A$

where

s_B - interindividual biological variation

s_A - analytical variation

k - factor dependent on number of measurements and confidence level



Exclusion of data Controls

- Quotients tested for bias:

- X/CAL, P/CAL, HIGH/LOW

- Quality goals

expressed with reference limits (H, L) and log transformed distributions according to

$$CV_B = [\ln(H) - \ln(L)]/4$$

and measurement uncertainty for the quotients, gives the following quality criteria:

- Optimum

$$B < 1/32 \cdot (\ln H - \ln L) + 2 \cdot (1/i + 1/j)^{1/2} \cdot CV_{CAL}$$

- Desirable

$$B < 2/32 \cdot (\ln H - \ln L) + 2 \cdot (1/i + 1/j)^{1/2} \cdot CV_{CAL}$$

- Minimum

$$B < 3/32 \cdot (\ln H - \ln L) + 2 \cdot (1/i + 1/j)^{1/2} \cdot CV_{CAL}$$



Exclusion of data Controls

- Result
 - No of exclusions
 - Effect on calculated reference limit
- Conclusion
 - Exclusion only used for sodium

Exclusion of data

Same person, different materials

- Absolute difference > 1.5 CVb and at least one outside reference limit
- No excluded: 1053 (0.8%)



Exclusion of data

Gross errors

- Many such errors corrected by report
- Rest (2-3) deleted



Exclusion of data Summary

- No of results: 126213
- No of results excluded
 - Person 1159 0.9%
 - Enzymes 15939 12.6%
 - Controls 2367 1.9%
 - Material 1053 0.8%
 - Gross errors 2 0.0%
 - Duplicate 1710 1.4%
- Not excluded: 104959



What have we NOT done?

- Frozen plasma
- Multivariate ref.intervals
- Influence of
 - all person parameters
 - sample tubes
 - geography